

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX092420\
 Data File : VX018576.D
 Acq On : 24 Sep 2020 16:39
 Operator : JC/SP
 Sample : L4118-12
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW203D-20200919

Quant Time: Sep 25 02:17:36 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X091720W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 17 12:17:36 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	432579	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	705100	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	699924	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	340414	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	318851	52.10	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.20%	
35) Dibromofluoromethane	5.47	113	251874	51.67	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.34%	
50) Toluene-d8	8.70	98	874479	49.38	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.76%	
62) 4-Bromofluorobenzene	11.12	95	326344	46.32	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.64%	
Target Compounds						
12) 1,1-Dichloroethene	2.36	96	6633	1.583	ug/l	90
24) 1,1-Dichloroethane	3.67	63	22986	2.612	ug/l	96
27) cis-1,2-Dichloroethene	4.57	96	15197	2.785	ug/l	89
30) Chloroform	5.18	83	6186	0.693	ug/l	97
32) 1,1,1-Trichloroethane	5.46	97	9686	1.156	ug/l #	45
44) Trichloroethene	7.19	130	96013	16.910	ug/l	95
64) Tetrachloroethene	9.32	164	3672	0.666	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX092420\
Data File : VX018576.D
Acq On : 24 Sep 2020 16:39
Operator : JC/SP
Sample : L4118-12
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW203D-20200919

Quant Time: Sep 25 02:17:36 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X091720W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 17 12:17:36 2020
Response via : Initial Calibration

