

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX092420\
 Data File : VX018578.D
 Acq On : 24 Sep 2020 17:24
 Operator : JC/SP
 Sample : L4118-14
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW201D1-20200919

Quant Time: Sep 25 02:22:42 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X091720W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 17 12:17:36 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	446172	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	714445	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	702691	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	340880	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	325961	51.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.28%	
35) Dibromofluoromethane	5.47	113	250127	50.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.28%	
50) Toluene-d8	8.70	98	891798	49.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.40%	
62) 4-Bromofluorobenzene	11.12	95	328191	45.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.96%	
Target Compounds						
7) Trichlorofluoromethane	1.92	101	2547	0.320	ug/l #	72
44) Trichloroethene	7.19	130	15582	2.709	ug/l	98
64) Tetrachloroethene	9.32	164	4277	0.773	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX092420\
Data File : VX018578.D
Acq On : 24 Sep 2020 17:24
Operator : JC/SP
Sample : L4118-14
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW201D1-20200919

Quant Time: Sep 25 02:22:42 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X091720W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 17 12:17:36 2020
Response via : Initial Calibration

