

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX092420\  
 Data File : VX018581.D  
 Acq On : 24 Sep 2020 18:31  
 Operator : JC/SP  
 Sample : L4118-17  
 Misc : 5.0mL/MSVOA X/WATER  
 ALS Vial : 24 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 TT180D-20200921

Quant Time: Sep 25 02:29:07 2020  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_X\METHOD\82X091720W.M  
 Quant Title : SW846 8260  
 QLast Update : Thu Sep 17 12:17:36 2020  
 Response via : Initial Calibration

| Internal Standards          | R.T.  | QIon | Response | Conc  | Units   | Dev(Min) |
|-----------------------------|-------|------|----------|-------|---------|----------|
| 1) Pentafluorobenzene       | 5.63  | 168  | 431248   | 50.00 | ug/l    | 0.00     |
| 34) 1,4-Difluorobenzene     | 6.84  | 114  | 705306   | 50.00 | ug/l    | 0.00     |
| 63) Chlorobenzene-d5        | 10.10 | 117  | 684916   | 50.00 | ug/l    | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 12.06 | 152  | 342935   | 50.00 | ug/l    | 0.00     |
| System Monitoring Compounds |       |      |          |       |         |          |
| 33) 1,2-Dichloroethane-d4   | 6.03  | 65   | 317892   | 52.11 | ug/l    | 0.00     |
| Spiked Amount 50.000        |       |      | Recovery | =     | 104.22% |          |
| 35) Dibromofluoromethane    | 5.47  | 113  | 248001   | 50.86 | ug/l    | 0.00     |
| Spiked Amount 50.000        |       |      | Recovery | =     | 101.72% |          |
| 50) Toluene-d8              | 8.70  | 98   | 873114   | 49.28 | ug/l    | 0.00     |
| Spiked Amount 50.000        |       |      | Recovery | =     | 98.56%  |          |
| 62) 4-Bromofluorobenzene    | 11.12 | 95   | 324457   | 46.04 | ug/l    | 0.00     |
| Spiked Amount 50.000        |       |      | Recovery | =     | 92.08%  |          |
| Target Compounds            |       |      |          |       |         |          |
| 24) 1,1-Dichloroethane      | 3.68  | 63   | 5594     | 0.638 | ug/l    | 96       |
| 44) Trichloroethene         | 7.19  | 130  | 12625    | 2.223 | ug/l    | 99       |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX092420\  
Data File : VX018581.D  
Acq On : 24 Sep 2020 18:31  
Operator : JC/SP  
Sample : L4118-17  
Misc : 5.0mL/MSVOA X/WATER  
ALS Vial : 24 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
TT180D-20200921

Quant Time: Sep 25 02:29:07 2020  
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA\_X\METHOD\82X091720W.M  
Quant Title : SW846 8260  
QLast Update : Thu Sep 17 12:17:36 2020  
Response via : Initial Calibration

