

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX092420\
 Data File : VX018586.D
 Acq On : 24 Sep 2020 20:23
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Sep 25 02:38:46 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X091720W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 17 12:17:36 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	74	0.00
2 T	Dichlorodifluoromethane	0.474	0.471	0.6	70	0.00
3 P	Chloromethane	0.555	0.495	10.8	69	0.00
4 C	Vinyl Chloride	0.588	0.539	8.3#	69	0.00
5 T	Bromomethane	0.361	0.367	-1.7	73	0.01
6 T	Chloroethane	0.336	0.312	7.1	69	0.00
7 T	Trichlorofluoromethane	0.893	0.946	-5.9	80	0.00
8 T	Diethyl Ether	0.313	0.296	5.4	74	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.480	0.478	0.4	74	0.00
10 T	Methyl Iodide	0.518	0.597	-15.3	75	0.00
11 T	Tert butyl alcohol	0.123	0.104	15.4	63	0.02
12 CM	1,1-Dichloroethene	0.484	0.472	2.5#	73	0.00
13 T	Acrolein	0.078	0.090	-15.4	88	0.00
14 T	Allyl chloride	0.938	0.822	12.4	66	0.00
15 T	Acrylonitrile	0.294	0.269	8.5	67	0.00
16 T	Acetone	0.271	0.232	14.4	66	0.01
17 T	Carbon Disulfide	1.446	1.311	9.3	69	0.00
18 T	Methyl Acetate	0.670	0.654	2.4	73	0.00
19 T	Methyl tert-butyl Ether	1.700	1.701	-0.1	73	0.00
20 T	Methylene Chloride	0.597	0.543	9.0	73	0.00
21 T	trans-1,2-Dichloroethene	0.563	0.521	7.5	72	0.00
22 T	Diisopropyl ether	1.764	1.699	3.7	71	0.00
23 T	Vinyl Acetate	1.563	1.488	4.8	69	0.00
24 P	1,1-Dichloroethane	1.017	0.959	5.7	72	0.00
25 T	2-Butanone	0.421	0.374	11.2	65	0.00
26 T	2,2-Dichloropropane	0.940	0.776	17.4	64	0.00
27 T	cis-1,2-Dichloroethene	0.631	0.578	8.4	70	0.00
28 T	Bromochloromethane	0.476	0.451	5.3	71	0.00
29 T	Tetrahydrofuran	0.266	0.238	10.5	65	0.00
30 C	Chloroform	1.032	1.039	-0.7#	76	0.00
31 T	Cyclohexane	0.856	0.784	8.4	68	0.00
32 T	1,1,1-Trichloroethane	0.968	0.977	-0.9	76	0.00
33 S	1,2-Dichloroethane-d4	0.707	0.673	4.8	76	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	72	0.00
35 S	Dibromofluoromethane	0.346	0.339	2.0	76	0.00
36 T	1,1-Dichloropropene	0.485	0.476	1.9	71	0.00
37 T	Ethyl Acetate	0.518	0.494	4.6	68	0.00
38 T	Carbon Tetrachloride	0.537	0.583	-8.6	77	0.00
39 T	Methylcyclohexane	0.545	0.507	7.0	67	0.00
40 TM	Benzene	1.391	1.342	3.5	71	0.00
41 T	Methacrylonitrile	0.288	0.270	6.2	69	0.00
42 TM	1,2-Dichloroethane	0.541	0.564	-4.3	76	0.00
43 T	Isopropyl Acetate	0.860	0.801	6.9	67	0.00
44 TM	Trichloroethene	0.403	0.402	0.2	73	0.00
45 C	1,2-Dichloropropane	0.382	0.366	4.2#	72	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\WX092420\
 Data File : VX018586.D
 Acq On : 24 Sep 2020 20:23
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Sep 25 02:38:46 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X091720W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 17 12:17:36 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.262	0.261	0.4	73	0.00
47 T	Bromodichloromethane	0.529	0.549	-3.8	75	0.00
48 T	Methyl methacrylate	0.419	0.398	5.0	67	0.00
49 T	1,4-Dioxane	0.007	0.005	28.6#	56	0.03
50 S	Toluene-d8	1.256	1.189	5.3	72	0.00
51 T	4-Methyl-2-Pentanone	0.512	0.499	2.5	68	0.00
52 CM	Toluene	0.875	0.864	1.3#	72	0.00
53 T	t-1,3-Dichloropropene	0.601	0.587	2.3	71	0.00
54 T	cis-1,3-Dichloropropene	0.619	0.601	2.9	69	0.00
55 T	1,1,2-Trichloroethane	0.363	0.362	0.3	72	0.00
56 T	Ethyl methacrylate	0.538	0.541	-0.6	70	0.00
57 T	1,3-Dichloropropane	0.610	0.607	0.5	72	0.00
58 T	2-Chloroethyl Vinyl ether	0.286	0.267	6.6	66	0.00
59 T	2-Hexanone	0.389	0.373	4.1	66	0.00
60 T	Dibromochloromethane	0.433	0.459	-6.0	76	0.00
61 T	1,2-Dibromoethane	0.386	0.401	-3.9	74	0.00
62 S	4-Bromofluorobenzene	0.500	0.463	7.4	71	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	74	0.00
64 T	Tetrachloroethene	0.394	0.380	3.6	74	0.00
65 PM	Chlorobenzene	1.030	0.998	3.1	73	0.00
66 T	1,1,1,2-Tetrachloroethane	0.413	0.411	0.5	74	0.00
67 C	Ethyl Benzene	1.782	1.743	2.2#	71	0.00
68 T	m/p-Xylenes	0.674	0.672	0.3	73	0.00
69 T	o-Xylene	0.638	0.622	2.5	71	0.00
70 T	Styrene	1.102	1.097	0.5	70	0.00
71 P	Bromoform	0.347	0.343	1.2	72	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	77	0.00
73 T	Isopropylbenzene	3.340	3.174	5.0	72	0.00
74 T	N-amyl acetate	1.504	1.316	12.5	65	0.00
75 P	1,1,2,2-Tetrachloroethane	1.106	1.009	8.8	71	0.00
76 T	1,2,3-Trichloropropane	1.051	0.924	12.1	68	0.00
77 T	Bromobenzene	0.922	0.828	10.2	71	0.00
78 T	n-propylbenzene	3.910	3.644	6.8	70	0.00
79 T	2-Chlorotoluene	2.367	2.163	8.6	70	0.00
80 T	1,3,5-Trimethylbenzene	2.857	2.738	4.2	72	0.00
81 T	trans-1,4-Dichloro-2-butene	0.410	0.344	16.1	64	0.00
82 T	4-Chlorotoluene	2.813	2.614	7.1	71	0.00
83 T	tert-Butylbenzene	2.774	2.630	5.2	71	0.00
84 T	1,2,4-Trimethylbenzene	2.877	2.755	4.2	71	0.00
85 T	sec-Butylbenzene	3.273	3.031	7.4	70	0.00
86 T	p-Isopropyltoluene	3.008	2.878	4.3	71	0.00
87 T	1,3-Dichlorobenzene	1.657	1.505	9.2	71	0.00
88 T	1,4-Dichlorobenzene	1.680	1.535	8.6	72	0.00
89 T	n-Butylbenzene	2.784	2.523	9.4	69	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX092420\
Data File : VX018586.D
Acq On : 24 Sep 2020 20:23
Operator : JC/SP
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Sep 25 02:38:46 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X091720W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 17 12:17:36 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.600	0.553	7.8	72	0.00
91 T	1,2-Dichlorobenzene	1.551	1.450	6.5	73	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.268	0.251	6.3	70	0.00
93 T	1,2,4-Trichlorobenzene	1.079	0.967	10.4	72	0.00
94 T	Hexachlorobutadiene	0.453	0.421	7.1	74	0.00
95 T	Naphthalene	3.260	2.999	8.0	68	0.00
96 T	1,2,3-Trichlorobenzene	1.032	0.942	8.7	70	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6