

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092424\
 Data File : VX043110.D
 Acq On : 24 Sep 2024 09:56
 Operator : JC/MD
 Sample : VX0924MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0924MBL01

Quant Time: Sep 25 01:20:25 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092324W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 24 03:13:52 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	90168	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	173018	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	163636	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	83634	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	77608	51.817	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 103.640%	
35) Dibromofluoromethane	5.385	113	60511	49.578	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 99.160%	
50) Toluene-d8	8.647	98	213907	51.802	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 103.600%	
62) 4-Bromofluorobenzene	11.079	95	86530	52.035	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 104.060%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092424\
Data File : VX043110.D
Acq On : 24 Sep 2024 09:56
Operator : JC/MD
Sample : VX0924MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0924MBL01

Quant Time: Sep 25 01:20:25 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092324W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 24 03:13:52 2024
Response via : Initial Calibration

