

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092424\
 Data File : VX043115.D
 Acq On : 24 Sep 2024 11:58
 Operator : JC/MD
 Sample : P4103-16ME
 Misc : 5.08g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WC-23ME

Quant Time: Sep 25 05:01:35 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092324W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 24 03:13:52 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	80428	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	158875	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	150231	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	81658	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	71597	53.593	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 107.180%	
35) Dibromofluoromethane	5.385	113	53099	47.378	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 94.760%	
50) Toluene-d8	8.647	98	194746	51.360	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 102.720%	
62) 4-Bromofluorobenzene	11.079	95	82645	54.123	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 108.240%	
Target Compounds						
					Qvalue	
39) Methylcyclohexane	7.373	83	1234	0.714	ug/l #	88
67) Ethyl Benzene	10.201	91	1527	0.272	ug/l	96
68) m/p-Xylenes	10.305	106	2129	1.020	ug/l	94
69) o-Xylene	10.646	106	621	0.297	ug/l	90
73) Isopropylbenzene	10.963	105	2526	0.412	ug/l	97
78) n-propylbenzene	11.305	91	19557	2.815	ug/l	99
80) 1,3,5-Trimethylbenzene	11.451	105	25700	5.072	ug/l	98
84) 1,2,4-Trimethylbenzene	11.750	105	93049	18.459	ug/l	99
85) sec-Butylbenzene	11.890	105	9206	1.435	ug/l	97
89) n-Butylbenzene	12.329	91	42566	9.168	ug/l #	54

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092424\
Data File : VX043115.D
Acq On : 24 Sep 2024 11:58
Operator : JC/MD
Sample : P4103-16ME
Misc : 5.08g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WC-23ME

Quant Time: Sep 25 05:01:35 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092324W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 24 03:13:52 2024
Response via : Initial Calibration

