

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092424\
 Data File : VX043118.D
 Acq On : 24 Sep 2024 13:07
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Sep 25 02:34:07 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092324W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 24 03:13:52 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	99262	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	195869	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	186760	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	96256	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	86810	52.651	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 105.300%	
35) Dibromofluoromethane	5.385	113	67062	48.535	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 97.080%	
50) Toluene-d8	8.647	98	240873	51.527	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 103.060%	
62) 4-Bromofluorobenzene	11.079	95	99568	52.890	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 105.780%	
Target Compounds						Qvalue
89) n-Butylbenzene	12.335	91	1476	0.270	ug/l	85
93) 1,2,4-Trichlorobenzene	13.591	180	560	0.290	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092424\
Data File : VX043118.D
Acq On : 24 Sep 2024 13:07
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Sep 25 02:34:07 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092324W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 24 03:13:52 2024
Response via : Initial Calibration

