

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092424\
 Data File : VX043130.D
 Acq On : 24 Sep 2024 18:08
 Operator : JC/MD
 Sample : P4164-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE134D2-20240919

Quant Time: Sep 25 01:34:10 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092324W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 24 03:13:52 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	92634	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	184298	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	177505	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	91681	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	82427	53.569	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 107.140%	
35) Dibromofluoromethane	5.385	113	62742	48.259	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 96.520%	
50) Toluene-d8	8.647	98	227663	51.759	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 103.520%	
62) 4-Bromofluorobenzene	11.079	95	95035	53.651	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 107.300%	
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.331	101	12108	11.797	ug/l	99
16) Acetone	2.379	43	1006	1.398	ug/l	93
40) Benzene	6.056	78	1474	0.313	ug/l #	83
44) Trichloroethene	7.123	130	102549	79.509	ug/l	97
49) 1,4-Dioxane	7.677	88	815	26.225	ug/l	92
52) Toluene	8.726	92	1144	0.377	ug/l	94
64) Tetrachloroethene	9.275	164	9586	7.453	ug/l	92
68) m/p-Xylenes	10.305	106	975	0.396	ug/l	76
69) o-Xylene	10.646	106	756	0.306	ug/l	98
84) 1,2,4-Trimethylbenzene	11.756	105	1519	0.268	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092424\
Data File : VX043130.D
Acq On : 24 Sep 2024 18:08
Operator : JC/MD
Sample : P4164-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE134D2-20240919

Quant Time: Sep 25 01:34:10 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092324W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 24 03:13:52 2024
Response via : Initial Calibration

