

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092425\
 Data File : VX047777.D
 Acq On : 24 Sep 2025 15:44
 Operator : JC/MD
 Sample : Q3097-23DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE134D4-20250912DL

Quant Time: Sep 25 01:17:08 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	133853	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	286287	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	290978	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	148581	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	112208	52.665	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 105.320%	
35) Dibromofluoromethane	5.373	113	91965	47.899	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 95.800%	
50) Toluene-d8	8.634	98	319132	48.036	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 96.080%	
62) 4-Bromofluorobenzene	11.061	95	137327	54.466	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 108.940%	
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.343	101	67631	43.684	ug/l	98
44) Trichloroethene	7.110	130	8324	4.041	ug/l	92
64) Tetrachloroethene	9.262	164	5711	3.011	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092425\
Data File : VX047777.D
Acq On : 24 Sep 2025 15:44
Operator : JC/MD
Sample : Q3097-23DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE134D4-20250912DL

Quant Time: Sep 25 01:17:08 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

