

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092425\
 Data File : VX047778.D
 Acq On : 24 Sep 2025 16:05
 Operator : JC/MD
 Sample : Q3097-25DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE134D3-20250912DL

Quant Time: Sep 25 01:17:32 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	133917	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	281574	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	283549	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	143934	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	109542	51.389	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 102.780%	
35) Dibromofluoromethane	5.379	113	90543	47.947	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 95.900%	
50) Toluene-d8	8.635	98	313894	48.039	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 96.080%	
62) 4-Bromofluorobenzene	11.067	95	133656	53.897	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 107.800%	
Target Compounds						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.337	101	104227	67.289	ug/l	99
44) Trichloroethene	7.117	130	35328	17.436	ug/l	90
64) Tetrachloroethene	9.256	164	39207	21.211	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092425\
Data File : VX047778.D
Acq On : 24 Sep 2025 16:05
Operator : JC/MD
Sample : Q3097-25DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE134D3-20250912DL

Quant Time: Sep 25 01:17:32 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

