

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092425\
 Data File : VX047781.D
 Acq On : 24 Sep 2025 17:07
 Operator : JC/MD
 Sample : Q3127-16DL 5X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TT101D1-20250915DL

Quant Time: Sep 25 01:18:46 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound		R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards							
1) Pentafluorobenzene		5.543	168	135889	50.000	ug/l	0.00
34) 1,4-Difluorobenzene		6.750	114	288957	50.000	ug/l	0.00
63) Chlorobenzene-d5		10.036	117	293313	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4		12.006	152	150005	50.000	ug/l	0.00
System Monitoring Compounds							
33) 1,2-Dichloroethane-d4		5.946	65	117408	54.280	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	108.560%	
35) Dibromofluoromethane		5.379	113	94316	48.669	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.340%	
50) Toluene-d8		8.634	98	326934	48.756	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.520%	
62) 4-Bromofluorobenzene		11.061	95	138184	54.299	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	108.600%	
Target Compounds							
							Qvalue
9) 1,1,2-Trichlorotrifluo...		2.337	101	3366	2.142	ug/l	96
44) Trichloroethene		7.110	130	70951	34.122	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092425\
Data File : VX047781.D
Acq On : 24 Sep 2025 17:07
Operator : JC/MD
Sample : Q3127-16DL 5X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TT101D1-20250915DL

Quant Time: Sep 25 01:18:46 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

