

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092425\
 Data File : VX047782.D
 Acq On : 24 Sep 2025 17:28
 Operator : JC/MD
 Sample : Q3127-06DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TT180D2-20250915DL

Quant Time: Sep 25 01:19:09 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.543	168	128490	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	270303	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	276987	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	135362	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	102904	50.314	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 100.620%	
35) Dibromofluoromethane	5.379	113	88190	48.649	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 97.300%	
50) Toluene-d8	8.634	98	306454	48.856	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 97.720%	
62) 4-Bromofluorobenzene	11.061	95	128982	54.181	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 108.360%	
Target Compounds						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.337	101	4595	3.092	ug/l	96
12) 1,1-Dichloroethene	2.325	96	2455	1.608	ug/l	96
44) Trichloroethene	7.110	130	71840	36.934	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092425\
Data File : VX047782.D
Acq On : 24 Sep 2025 17:28
Operator : JC/MD
Sample : Q3127-06DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TT180D2-20250915DL

Quant Time: Sep 25 01:19:09 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

