

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092425\
 Data File : VX047786.D
 Acq On : 24 Sep 2025 18:51
 Operator : JC/MD
 Sample : Q3127-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE135D2-20250915

Quant Time: Sep 25 01:20:42 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	128068	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	276429	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	279711	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	142150	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	110362	54.138	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 108.280%	
35) Dibromofluoromethane	5.373	113	91361	49.281	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 98.560%	
50) Toluene-d8	8.635	98	312620	48.734	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 97.460%	
62) 4-Bromofluorobenzene	11.061	95	133284	54.747	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 109.500%	
Target Compounds						Qvalue
16) Acetone	2.380	43	3415	3.853	ug/l #	85
45) 1,2-Dichloropropane	7.422	63	6397	3.036	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092425\
Data File : VX047786.D
Acq On : 24 Sep 2025 18:51
Operator : JC/MD
Sample : Q3127-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE135D2-20250915

Quant Time: Sep 25 01:20:42 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

