

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092425\
 Data File : VX047788.D
 Acq On : 24 Sep 2025 19:33
 Operator : JC/MD
 Sample : Q3127-18
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BPOW1-8-20250917

Quant Time: Sep 25 01:21:40 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound		R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards							
1) Pentafluorobenzene		5.544	168	128659	50.000	ug/l	0.00
34) 1,4-Difluorobenzene		6.751	114	271871	50.000	ug/l	0.00
63) Chlorobenzene-d5		10.037	117	278272	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4		12.006	152	143672	50.000	ug/l	0.00
System Monitoring Compounds							
33) 1,2-Dichloroethane-d4		5.946	65	108638	53.048	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	106.100%	
35) Dibromofluoromethane		5.379	113	89155	48.897	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.800%	
50) Toluene-d8		8.635	98	306192	48.532	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.060%	
62) 4-Bromofluorobenzene		11.061	95	131657	54.986	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	109.980%	
Target Compounds							
16) Acetone		2.380	43	1529	1.717	ug/l #	88
44) Trichloroethene		7.123	130	749	0.383	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092425\
Data File : VX047788.D
Acq On : 24 Sep 2025 19:33
Operator : JC/MD
Sample : Q3127-18
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BPOW1-8-20250917

Quant Time: Sep 25 01:21:40 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

