

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX092520\
 Data File : VX018591.D
 Acq On : 25 Sep 2020 11:08
 Operator : JC/SP
 Sample : VX0925WBL02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0925WBL02

Quant Time: Sep 25 18:27:26 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X091720W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 17 12:17:36 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	460305	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	710753	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	699545	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	328285	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	320795	49.26	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.52%	
35) Dibromofluoromethane	5.46	113	254068	51.71	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.42%	
50) Toluene-d8	8.70	98	877311	49.14	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.28%	
62) 4-Bromofluorobenzene	11.12	95	326873	46.03	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.06%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX092520\
Data File : VX018591.D
Acq On : 25 Sep 2020 11:08
Operator : JC/SP
Sample : VX0925WBL02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampled :
VX0925WBL02

Quant Time: Sep 25 18:27:26 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X091720W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 17 12:17:36 2020
Response via : Initial Calibration

