

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX092520\
 Data File : VX018604.D
 Acq On : 25 Sep 2020 18:32
 Operator : JC/SP
 Sample : L4140-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-1

Quant Time: Sep 28 09:34:31 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X091720W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 17 12:17:36 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	457597	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	718297	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	727436	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	365389	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	327158	50.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.08%	
35) Dibromofluoromethane	5.47	113	252222	50.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.58%	
50) Toluene-d8	8.70	98	882003	48.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.78%	
62) 4-Bromofluorobenzene	11.12	95	348833	48.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.22%	
Target Compounds						
16) Acetone	2.43	43	7985	3.225	ug/l	Qvalue 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX092520\
Data File : VX018604.D
Acq On : 25 Sep 2020 18:32
Operator : JC/SP
Sample : L4140-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-1

Quant Time: Sep 28 09:34:31 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X091720W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 17 12:17:36 2020
Response via : Initial Calibration

