

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX092520\
 Data File : VX018606.D
 Acq On : 25 Sep 2020 19:17
 Operator : JC/SP
 Sample : L4140-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-4

Quant Time: Sep 28 09:37:36 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X091720W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 17 12:17:36 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	455464	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	717038	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	718287	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	359887	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	325202	50.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.94%	
35) Dibromofluoromethane	5.47	113	253909	51.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.44%	
50) Toluene-d8	8.70	98	890062	49.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.84%	
62) 4-Bromofluorobenzene	11.12	95	347887	48.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.12%	
Target Compounds						
19) Methyl tert-butyl Ether	3.17	73	41423	2.675	ug/l	Qvalue 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX092520\
Data File : VX018606.D
Acq On : 25 Sep 2020 19:17
Operator : JC/SP
Sample : L4140-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-4

Quant Time: Sep 28 09:37:36 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X091720W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 17 12:17:36 2020
Response via : Initial Calibration

