

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX092520\
 Data File : VX018607.D
 Acq On : 25 Sep 2020 19:39
 Operator : JC/SP
 Sample : L4140-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-2

Quant Time: Sep 28 09:38:52 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X091720W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 17 12:17:36 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	440348	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	692571	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	684519	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	340218	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	316678	50.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.68%	
35) Dibromofluoromethane	5.47	113	240270	50.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.36%	
50) Toluene-d8	8.70	98	849160	48.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.62%	
62) 4-Bromofluorobenzene	11.12	95	323171	46.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.40%	
Target Compounds						
					Qvalue	
4) Vinyl Chloride	1.40	62	3739	0.723	ug/l #	70
16) Acetone	2.43	43	6156	2.584	ug/l #	88
19) Methyl tert-butyl Ether	3.17	73	6590	0.440	ug/l #	82
65) Chlorobenzene	10.12	112	8701	0.617	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX092520\
Data File : VX018607.D
Acq On : 25 Sep 2020 19:39
Operator : JC/SP
Sample : L4140-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-2

Quant Time: Sep 28 09:38:52 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X091720W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 17 12:17:36 2020
Response via : Initial Calibration

