

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092523\
 Data File : VX037887.D
 Acq On : 25 Sep 2023 11:23
 Operator : JC/MD
 Sample : 04595-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-RW10A-GW-460-462

Quant Time: Sep 26 06:56:50 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091523W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 16 04:55:44 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	269296	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	452436	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	427981	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	220880	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	164642	46.534	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	93.060%
35) Dibromofluoromethane	5.391	113	144204	47.635	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.260%
50) Toluene-d8	8.647	98	540340	48.038	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.080%
62) 4-Bromofluorobenzene	11.079	95	204754	50.094	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	100.180%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	7272	4.272	ug/l	93
17) Carbon Disulfide	2.514	76	4601	0.698	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092523\
Data File : VX037887.D
Acq On : 25 Sep 2023 11:23
Operator : JC/MD
Sample : 04595-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-RW10A-GW-460-462

Quant Time: Sep 26 06:56:50 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091523W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 16 04:55:44 2023
Response via : Initial Calibration

