

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092523\
 Data File : VX037894.D
 Acq On : 25 Sep 2023 14:06
 Operator : JC/MD
 Sample : PB155756TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB155756TB

Quant Time: Sep 26 06:58:00 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091523W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 16 04:55:44 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	224951	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	387130	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	384922	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	193736	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	142949	48.367	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	96.740%
35) Dibromofluoromethane	5.391	113	121355	46.850	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.700%
50) Toluene-d8	8.653	98	482797	50.163	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.320%
62) 4-Bromofluorobenzene	11.079	95	180981	51.747	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	103.500%
Target Compounds						
16) Acetone	2.380	43	40052	28.170	ug/l	Qvalue 92
20) Methylene Chloride	2.794	84	12498	4.732	ug/l	# 82
43) Isopropyl Acetate	6.342	43	145487	23.194	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092523\
Data File : VX037894.D
Acq On : 25 Sep 2023 14:06
Operator : JC/MD
Sample : PB155756TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB155756TB

Quant Time: Sep 26 06:58:00 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091523W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 16 04:55:44 2023
Response via : Initial Calibration

