

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092523\
 Data File : VX037895.D
 Acq On : 25 Sep 2023 14:44
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Sep 26 06:58:12 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091523W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 16 04:55:44 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	222421	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	383181	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	371778	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	188620	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	140585	48.108	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	96.220%
35) Dibromofluoromethane	5.385	113	117945	46.002	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.000%
50) Toluene-d8	8.647	98	472866	49.637	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.280%
62) 4-Bromofluorobenzene	11.079	95	174855	50.511	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	101.020%
Target Compounds						
					Qvalue	
20) Methylene Chloride	2.788	84	1263	0.461	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092523\
Data File : VX037895.D
Acq On : 25 Sep 2023 14:44
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Sep 26 06:58:12 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091523W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 16 04:55:44 2023
Response via : Initial Calibration

