

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092524\
 Data File : VX043137.D
 Acq On : 25 Sep 2024 08:58
 Operator : JC/MD
 Sample : VX0925MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0925MBL01

Quant Time: Sep 26 01:27:59 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092324W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 24 03:13:52 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	87968	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	174591	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	162676	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	83304	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	77619	53.120	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	106.240%
35) Dibromofluoromethane	5.385	113	58759	47.709	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.420%
50) Toluene-d8	8.647	98	213486	51.234	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	102.460%
62) 4-Bromofluorobenzene	11.079	95	86450	51.518	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	103.040%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092524\
Data File : VX043137.D
Acq On : 25 Sep 2024 08:58
Operator : JC/MD
Sample : VX0925MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0925MBL01

Quant Time: Sep 26 01:27:59 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092324W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 24 03:13:52 2024
Response via : Initial Calibration

