

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092524\
 Data File : VX043138.D
 Acq On : 25 Sep 2024 09:21
 Operator : JC/MD
 Sample : VX0925WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0925WBL01

Quant Time: Sep 26 01:28:36 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092324W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 24 03:13:52 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	91571	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	182684	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	171642	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	88453	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	82155	54.012	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	108.020%
35) Dibromofluoromethane	5.385	113	62498	48.496	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.000%
50) Toluene-d8	8.647	98	221363	50.771	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.540%
62) 4-Bromofluorobenzene	11.079	95	91013	51.835	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	103.660%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092524\
Data File : VX043138.D
Acq On : 25 Sep 2024 09:21
Operator : JC/MD
Sample : VX0925WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0925WBL01

Quant Time: Sep 26 01:28:36 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092324W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 24 03:13:52 2024
Response via : Initial Calibration

