

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092524\
 Data File : VX043141.D
 Acq On : 25 Sep 2024 10:34
 Operator : JC/MD
 Sample : P4164-09DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE134D3-20240919DL

Quant Time: Sep 26 01:31:01 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092324W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 24 03:13:52 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	88750	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	175473	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	161763	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	79081	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	77544	52.601	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	105.200%
35) Dibromofluoromethane	5.385	113	59693	48.223	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.440%
50) Toluene-d8	8.647	98	211837	50.583	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.160%
62) 4-Bromofluorobenzene	11.079	95	85146	50.486	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	100.980%
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.331	101	98655	100.325	ug/l	99
44) Trichloroethene	7.129	130	26124	21.273	ug/l	94
64) Tetrachloroethene	9.275	164	32348	27.600	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092524\
Data File : VX043141.D
Acq On : 25 Sep 2024 10:34
Operator : JC/MD
Sample : P4164-09DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE134D3-20240919DL

Quant Time: Sep 26 01:31:01 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092324W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 24 03:13:52 2024
Response via : Initial Calibration

