

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092524\
 Data File : VX043150.D
 Acq On : 25 Sep 2024 14:02
 Operator : JC/MD
 Sample : P4164-24
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FB

Quant Time: Sep 26 01:35:06 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092324W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 24 03:13:52 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	87323	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	173643	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	169695	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	87509	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	79570	54.858	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 109.720%	
35) Dibromofluoromethane	5.391	113	59405	48.496	ug/l	0.01
Spiked Amount	50.000	Range	75 - 124	Recovery	= 97.000%	
50) Toluene-d8	8.647	98	216326	52.199	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 104.400%	
62) 4-Bromofluorobenzene	11.079	95	89352	53.538	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 107.080%	
Target Compounds						
					Qvalue	
95) Naphthalene	13.774	128	8836	1.449	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092524\
Data File : VX043150.D
Acq On : 25 Sep 2024 14:02
Operator : JC/MD
Sample : P4164-24
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FB

Quant Time: Sep 26 01:35:06 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092324W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 24 03:13:52 2024
Response via : Initial Calibration

