

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092525\
 Data File : VX047800.D
 Acq On : 25 Sep 2025 11:03
 Operator : JC/MD
 Sample : Q3165-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VPB184-HYD-20250919

Quant Time: Sep 26 01:24:28 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.537	168	120300	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.745	114	251100	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	252620	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	124631	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	95225	49.729	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	99.460%
35) Dibromofluoromethane	5.373	113	79370	47.131	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.260%
50) Toluene-d8	8.634	98	278502	47.795	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.580%
62) 4-Bromofluorobenzene	11.061	95	116832	52.830	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	105.660%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092525\
Data File : VX047800.D
Acq On : 25 Sep 2025 11:03
Operator : JC/MD
Sample : Q3165-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VPB184-HYD-20250919

Quant Time: Sep 26 01:24:28 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

