

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092525\
 Data File : VX047805.D
 Acq On : 25 Sep 2025 12:49
 Operator : JC/MD
 Sample : Q3165-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-184-DUP-20250922

Quant Time: Sep 26 01:26:36 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.543	168	118043	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.744	114	252631	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.036	117	261796	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	135087	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	102254	54.421	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 108.840%	
35) Dibromofluoromethane	5.373	113	84610	49.939	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 99.880%	
50) Toluene-d8	8.634	98	288250	49.168	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 98.340%	
62) 4-Bromofluorobenzene	11.067	95	123534	55.522	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 111.040%	
Target Compounds						
						Qvalue
3) Chloromethane	1.306	50	1013	0.631	ug/l	99
16) Acetone	2.379	43	7787	9.532	ug/l	95
17) Carbon Disulfide	2.519	76	3075	0.801	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092525\
Data File : VX047805.D
Acq On : 25 Sep 2025 12:49
Operator : JC/MD
Sample : Q3165-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-184-DUP-20250922

Quant Time: Sep 26 01:26:36 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

