

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092525\
 Data File : VX047807.D
 Acq On : 25 Sep 2025 13:31
 Operator : JC/MD
 Sample : Q3127-04DL 5X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP07-20250915DL

Quant Time: Sep 26 01:27:26 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	121002	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.745	114	255503	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	261171	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	129528	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	98044	50.904	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	101.800%
35) Dibromofluoromethane	5.379	113	83769	48.886	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.780%
50) Toluene-d8	8.634	98	290222	48.948	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.900%
62) 4-Bromofluorobenzene	11.067	95	121188	53.856	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	107.720%
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.331	101	760	0.543	ug/l	96
27) cis-1,2-Dichloroethene	4.483	96	1119	0.590	ug/l	90
44) Trichloroethene	7.110	130	74427	40.480	ug/l	96
64) Tetrachloroethene	9.256	164	4448	2.613	ug/l	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092525\
Data File : VX047807.D
Acq On : 25 Sep 2025 13:31
Operator : JC/MD
Sample : Q3127-04DL 5X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP07-20250915DL

Quant Time: Sep 26 01:27:26 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

