

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092525\
 Data File : VX047809.D
 Acq On : 25 Sep 2025 14:14
 Operator : JC/MD
 Sample : Q3127-13
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TT119S1-20250916

Quant Time: Sep 26 01:28:16 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.543	168	122798	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	255684	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	260354	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	131454	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	98318	50.300	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 100.600%	
35) Dibromofluoromethane	5.373	113	83213	48.528	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 97.060%	
50) Toluene-d8	8.634	98	287424	48.442	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 96.880%	
62) 4-Bromofluorobenzene	11.067	95	119966	53.275	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 106.540%	
Target Compounds						
					Qvalue	
44) Trichloroethene	7.110	130	2651	1.441	ug/l	77
64) Tetrachloroethene	9.262	164	658	0.388	ug/l	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092525\
Data File : VX047809.D
Acq On : 25 Sep 2025 14:14
Operator : JC/MD
Sample : Q3127-13
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TT119S1-20250916

Quant Time: Sep 26 01:28:16 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

