

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092525\
 Data File : VX047812.D
 Acq On : 25 Sep 2025 15:17
 Operator : JC/MD
 Sample : Q3128-09
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BPOW6-11-20250916

Quant Time: Sep 26 01:29:30 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	116305	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	248243	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	257163	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	128330	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	98484	53.198	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	106.400%
35) Dibromofluoromethane	5.379	113	82203	49.376	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.760%
50) Toluene-d8	8.634	98	282956	49.118	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.240%
62) 4-Bromofluorobenzene	11.067	95	119034	54.445	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	108.900%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092525\
Data File : VX047812.D
Acq On : 25 Sep 2025 15:17
Operator : JC/MD
Sample : Q3128-09
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BPOW6-11-20250916

Quant Time: Sep 26 01:29:30 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

