

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092525\
 Data File : VX047815.D
 Acq On : 25 Sep 2025 16:21
 Operator : JC/MD
 Sample : Q3128-12
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BPOW6-6-20250917

Quant Time: Sep 26 01:30:39 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.537	168	127711	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.745	114	269892	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	279053	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	143881	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	111634	54.915	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 109.840%			
35) Dibromofluoromethane	5.373	113	90854	50.194	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 100.380%			
50) Toluene-d8	8.634	98	310839	49.630	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 99.260%			
62) 4-Bromofluorobenzene	11.067	95	133248	56.058	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 112.120%			
Target Compounds						Qvalue
68) m/p-Xylenes	10.287	106	1337	0.328	ug/l	91
69) o-Xylene	10.628	106	994	0.252	ug/l	78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092525\
Data File : VX047815.D
Acq On : 25 Sep 2025 16:21
Operator : JC/MD
Sample : Q3128-12
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BPOW6-6-20250917

Quant Time: Sep 26 01:30:39 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

