

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092525\
 Data File : VX047817.D
 Acq On : 25 Sep 2025 17:03
 Operator : JC/MD
 Sample : Q3127-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP08-20250915

Quant Time: Sep 26 01:31:36 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	113877	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	245918	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	250155	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	122323	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	99112	54.678	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 109.360%	
35) Dibromofluoromethane	5.373	113	85257	51.694	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 103.380%	
50) Toluene-d8	8.635	98	287005	50.292	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 100.580%	
62) 4-Bromofluorobenzene	11.061	95	120370	55.577	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 111.160%	
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.173	85	679	0.576	ug/l	85
9) 1,1,2-Trichlorotrifluo...	2.331	101	7257	5.510	ug/l	98
12) 1,1-Dichloroethene	2.325	96	2185	1.615	ug/l	89
16) Acetone	2.386	43	2486	3.154	ug/l #	80
24) 1,1-Dichloroethane	3.617	63	2779	0.966	ug/l #	91
27) cis-1,2-Dichloroethene	4.489	96	5971	3.347	ug/l	91
30) Chloroform	5.099	83	1488	0.512	ug/l #	83
44) Trichloroethene	7.111	130	138235	78.115	ug/l	97
64) Tetrachloroethene	9.263	164	861	0.528	ug/l #	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092525\
Data File : VX047817.D
Acq On : 25 Sep 2025 17:03
Operator : JC/MD
Sample : Q3127-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP08-20250915

Quant Time: Sep 26 01:31:36 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

