

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092525\
 Data File : VX047826.D
 Acq On : 26 Sep 2025 02:10
 Operator : JC/MD
 Sample : VX0925WBL02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0925WBL02

Quant Time: Sep 26 03:56:40 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.538	168	121983	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.745	114	259174	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	269510	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	138899	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	105522	54.346	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 108.700%	
35) Dibromofluoromethane	5.373	113	84777	48.774	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 97.540%	
50) Toluene-d8	8.635	98	292890	48.698	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 97.400%	
62) 4-Bromofluorobenzene	11.067	95	125101	54.807	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 109.620%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092525\
Data File : VX047826.D
Acq On : 26 Sep 2025 02:10
Operator : JC/MD
Sample : VX0925WBL02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0925WBL02

Quant Time: Sep 26 03:56:40 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

