

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092525\
 Data File : VX047829.D
 Acq On : 26 Sep 2025 03:31
 Operator : JC/MD
 Sample : Q3156-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE10-MW01S-20250918

Quant Time: Sep 26 06:52:39 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.532	168	123505	50.000	ug/l	-0.01
34) 1,4-Difluorobenzene	6.739	114	257582	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	255230	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	123234	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.934	65	96094	48.881	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.760%
35) Dibromofluoromethane	5.367	113	83114	48.113	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.220%
50) Toluene-d8	8.629	98	285991	47.845	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.700%
62) 4-Bromofluorobenzene	11.061	95	117321	51.716	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	103.440%
Target Compounds						
					Qvalue	
16) Acetone	2.374	43	1853	2.168	ug/l	97
64) Tetrachloroethene	9.257	164	9922	5.963	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092525\
Data File : VX047829.D
Acq On : 26 Sep 2025 03:31
Operator : JC/MD
Sample : Q3156-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE10-MW01S-20250918

Quant Time: Sep 26 06:52:39 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

