

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092525\
 Data File : VX047836.D
 Acq On : 26 Sep 2025 05:58
 Operator : JC/MD
 Sample : Q3156-09
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 48 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE132D3-20250918

Quant Time: Sep 26 06:57:20 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.538	168	122520	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	255492	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	257823	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	126419	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	99183	50.858	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	101.720%
35) Dibromofluoromethane	5.373	113	82450	48.119	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.240%
50) Toluene-d8	8.635	98	284527	47.989	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.980%
62) 4-Bromofluorobenzene	11.067	95	117745	52.328	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	104.660%
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.179	85	1073	0.846	ug/l	89
9) 1,1,2-Trichlorotrifluo...	2.343	101	5346	3.772	ug/l	97
27) cis-1,2-Dichloroethene	4.477	96	1465	0.763	ug/l #	67
44) Trichloroethene	7.117	130	129214	70.282	ug/l	93
64) Tetrachloroethene	9.263	164	3225	1.919	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092525\
Data File : VX047836.D
Acq On : 26 Sep 2025 05:58
Operator : JC/MD
Sample : Q3156-09
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 48 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE132D3-20250918

Quant Time: Sep 26 06:57:20 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

