

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX092618\
 Data File : VX004745.D
 Acq On : 27 Sep 2018 01:31
 Operator : JC/MD
 Sample : J5029-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FB-092018-02

Quant Time: Sep 27 06:28:15 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 26 14:17:02 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	270047	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	391897	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	381533	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	216793	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	204976	53.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.28%	
35) Dibromofluoromethane	5.50	113	175592	52.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.74%	
50) Toluene-d8	8.71	98	567839	51.43	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.86%	
62) 4-Bromofluorobenzene	11.14	95	208310	52.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.74%	
Target Compounds						
16) Acetone	2.45	43	75376	33.468	ug/l	97
51) 4-Methyl-2-Pentanone	8.65	43	9478	1.734	ug/l	92
95) Naphthalene	13.83	128	47020	3.289	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX092618\
Data File : VX004745.D
Acq On : 27 Sep 2018 01:31
Operator : JC/MD
Sample : J5029-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampled :
FB-092018-02

Quant Time: Sep 27 06:28:15 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 26 14:17:02 2018
Response via : Initial Calibration

