

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX092618\
 Data File : VX004748.D
 Acq On : 27 Sep 2018 02:51
 Operator : JC/MD
 Sample : J5029-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 T11-MW-7-9.0-092018

Quant Time: Sep 27 06:38:22 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 26 14:17:02 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	290110	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	428698	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	403309	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	228820	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	208229	50.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.50%	
35) Dibromofluoromethane	5.51	113	180486	49.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.36%	
50) Toluene-d8	8.72	98	606394	50.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.40%	
62) 4-Bromofluorobenzene	11.14	95	221027	50.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.58%	
Target Compounds						
16) Acetone	2.45	43	3398	1.404	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX092618\
Data File : VX004748.D
Acq On : 27 Sep 2018 02:51
Operator : JC/MD
Sample : J5029-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 31 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
T11-MW-7-9.0-092018

Quant Time: Sep 27 06:38:22 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 26 14:17:02 2018
Response via : Initial Calibration

