

Data File : VX004749.D
Acq On : 27 Sep 2018 03:18
Operator : JC/MD
Sample : J5029-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
T11-MW-2-8.0-092018

Quant Time: Oct 04 04:50:07 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 26 14:17:02 2018
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	305375	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	433133	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	406030	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	231897	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	213630	48.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.94%	
35) Dibromofluoromethane	5.51	113	187351	51.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.08%	
50) Toluene-d8	8.71	98	605900	49.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.30%	
62) 4-Bromofluorobenzene	11.14	95	223015	50.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.46%	
Target Compounds						
11) Tert butyl alcohol	3.07	59	11086	9.089	ug/l #	93
16) Acetone	2.45	43	26947	10.581	ug/l	90
25) 2-Butanone	4.70	43	5700	1.518	ug/l	95
95) Naphthalene	13.83	128	8161	1.136	ug/l #	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VX004749.D
Acq On : 27 Sep 2018 03:18
Operator : JC/MD
Sample : J5029-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
T11-MW-2-8.0-092018

Quant Time: Oct 04 04:50:07 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 26 14:17:02 2018
Response via : Initial Calibration

