

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX092619\
 Data File : VX012671.D
 Acq On : 26 Sep 2019 14:06
 Operator : JC/SP
 Sample : K5049-09
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TRIP-BLANK

Quant Time: Sep 27 00:29:42 2019
 Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X092319W.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 23 12:27:01 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	168227	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	290083	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	252784	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	101073	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	122173	49.87	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.74%	
35) Dibromofluoromethane	5.48	113	88074	50.98	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.96%	
50) Toluene-d8	8.71	98	349207	52.49	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.98%	
62) 4-Bromofluorobenzene	11.14	95	126771	48.46	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.92%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX092619\
Data File : VX012671.D
Acq On : 26 Sep 2019 14:06
Operator : JC/SP
Sample : K5049-09
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TRIP-BLANK

Quant Time: Sep 27 00:29:42 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X092319W.M
Quant Title : SW846 8260
QLast Update : Mon Sep 23 12:27:01 2019
Response via : Initial Calibration

