

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX092619\
 Data File : VX012676.D
 Acq On : 26 Sep 2019 16:03
 Operator : JC/SP
 Sample : K5048-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 S49-0528

Quant Time: Sep 27 00:54:49 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092319W.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 23 12:27:01 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	175094	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	305495	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	260951	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	102203	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	133589	52.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.78%	
35) Dibromofluoromethane	5.49	113	95288	52.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.74%	
50) Toluene-d8	8.71	98	372909	53.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.44%	
62) 4-Bromofluorobenzene	11.14	95	130536	47.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.76%	
Target Compounds						
4) Vinyl Chloride	1.40	62	1209	0.538	ug/l	92
16) Acetone	2.43	43	3731	2.763	ug/l #	78
21) trans-1,2-Dichloroethene	3.16	96	8081	3.875	ug/l	96
27) cis-1,2-Dichloroethene	4.59	96	2994	1.237	ug/l	92
44) Trichloroethene	7.20	130	1440	0.650	ug/l	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX092619\
Data File : VX012676.D
Acq On : 26 Sep 2019 16:03
Operator : JC/SP
Sample : K5048-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
S49-0528

Quant Time: Sep 27 00:54:49 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X092319W.M
Quant Title : SW846 8260
QLast Update : Mon Sep 23 12:27:01 2019
Response via : Initial Calibration

