

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX092619\
 Data File : VX012683.D
 Acq On : 26 Sep 2019 18:47
 Operator : JC/SP
 Sample : K5049-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-2R

Quant Time: Sep 27 01:32:12 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092319W.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 23 12:27:01 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	160893	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	279160	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	225759	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	80643	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	117703	50.24	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.48%	
35) Dibromofluoromethane	5.49	113	84310	50.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.42%	
50) Toluene-d8	8.71	98	328114	51.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.50%	
62) 4-Bromofluorobenzene	11.14	95	106259	42.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	84.42%	
Target Compounds						
11) Tert butyl alcohol	3.03	59	167453	286.110	ug/l	100
16) Acetone	2.43	43	7414	5.974	ug/l #	88
19) Methyl tert-butyl Ether	3.19	73	116054	17.724	ug/l	97
85) sec-Butylbenzene	11.94	105	7842	1.269	ug/l #	69

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX092619\
Data File : VX012683.D
Acq On : 26 Sep 2019 18:47
Operator : JC/SP
Sample : K5049-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-2R

Quant Time: Sep 27 01:32:12 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X092319W.M
Quant Title : SW846 8260
QLast Update : Mon Sep 23 12:27:01 2019
Response via : Initial Calibration

