

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092623\
 Data File : VX037909.D
 Acq On : 26 Sep 2023 16:02
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Sep 27 01:16:14 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091523W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 16 04:55:44 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	90	0.00
2 T	Dichlorodifluoromethane	0.527	0.439	16.7	76	0.00
3 P	Chloromethane	0.573	0.464	19.0	79	0.00
4 C	Vinyl Chloride	0.619	0.531	14.2#	78	0.00
5 T	Bromomethane	0.443	0.418	5.6	84	0.00
6 T	Chloroethane	0.426	0.367	13.8	81	0.00
7 T	Trichlorofluoromethane	0.995	0.941	5.4	84	0.00
8 T	Diethyl Ether	0.385	0.351	8.8	85	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.505	0.480	5.0	86	0.00
10 T	Methyl Iodide	0.668	0.583	12.7	76	0.00
11 T	Tert butyl alcohol	0.141	0.116	17.7	77	-0.01
12 CM	1,1-Dichloroethene	0.488	0.442	9.4#	81	0.00
13 T	Acrolein	0.139	0.128	7.9	85	0.00
14 T	Allyl chloride	0.812	0.726	10.6	81	0.00
15 T	Acrylonitrile	0.314	0.281	10.5	81	0.00
16 T	Acetone	0.316	0.268	15.2	83	0.00
17 T	Carbon Disulfide	1.223	0.972	20.5	72	0.00
18 T	Methyl Acetate	1.140	1.043	8.5	84	0.00
19 T	Methyl tert-butyl Ether	1.794	1.687	6.0	86	-0.01
20 T	Methylene Chloride	0.641	0.549	14.4	85	0.00
21 T	trans-1,2-Dichloroethene	0.542	0.505	6.8	84	0.00
22 T	Diisopropyl ether	1.692	1.606	5.1	86	0.00
23 T	Vinyl Acetate	1.245	1.182	5.1	84	-0.01
24 P	1,1-Dichloroethane	1.014	0.941	7.2	85	0.00
25 T	2-Butanone	0.446	0.385	13.7	79	-0.02
26 T	2,2-Dichloropropane	0.819	0.796	2.8	89	0.00
27 T	cis-1,2-Dichloroethene	0.664	0.623	6.2	85	0.00
28 T	Bromochloromethane	0.473	0.429	9.3	85	-0.01
29 T	Tetrahydrofuran	0.277	0.240	13.4	78	-0.02
30 C	Chloroform	1.092	1.047	4.1#	87	0.00
31 T	Cyclohexane	0.843	0.763	9.5	80	0.00
32 T	1,1,1-Trichloroethane	0.925	0.905	2.2	86	0.00
33 S	1,2-Dichloroethane-d4	0.657	0.624	5.0	85	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	88	0.00
35 S	Dibromofluoromethane	0.335	0.332	0.9	86	0.00
36 T	1,1-Dichloropropene	0.461	0.441	4.3	84	0.00
37 T	Ethyl Acetate	0.507	0.461	9.1	80	-0.02
38 T	Carbon Tetrachloride	0.482	0.476	1.2	85	0.00
39 T	Methylcyclohexane	0.585	0.549	6.2	82	0.00
40 TM	Benzene	1.417	1.333	5.9	83	-0.01
41 T	Methacrylonitrile	0.258	0.245	5.0	84	-0.03
42 TM	1,2-Dichloroethane	0.512	0.491	4.1	86	-0.02
43 T	Isopropyl Acetate	0.810	0.761	6.0	82	0.00
44 TM	Trichloroethene	0.395	0.367	7.1	86	-0.01
45 C	1,2-Dichloropropane	0.368	0.357	3.0#	87	0.00
46 T	Dibromomethane	0.265	0.262	1.1	87	0.00
47 T	Bromodichloromethane	0.488	0.501	-2.7	88	0.00
48 T	Methyl methacrylate	0.387	0.363	6.2	81	-0.01

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092623\
 Data File : VX037909.D
 Acq On : 26 Sep 2023 16:02
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Sep 27 01:16:14 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091523W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 16 04:55:44 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.010	0.008	20.0	75	-0.02
50 S	Toluene-d8	1.243	1.243	0.0	86	0.00
51 T	4-Methyl-2-Pentanone	0.514	0.477	7.2	81	0.00
52 CM	Toluene	0.913	0.880	3.6#	84	0.00
53 T	t-1,3-Dichloropropene	0.510	0.541	-6.1	88	-0.01
54 T	cis-1,3-Dichloropropene	0.564	0.578	-2.5	87	0.00
55 T	1,1,2-Trichloroethane	0.385	0.386	-0.3	87	0.00
56 T	Ethyl methacrylate	0.575	0.564	1.9	84	0.00
57 T	1,3-Dichloropropane	0.623	0.604	3.0	86	0.00
58 T	2-Chloroethyl Vinyl ether	0.268	0.243	9.3	81	0.00
59 T	2-Hexanone	0.400	0.370	7.5	81	0.00
60 T	Dibromochloromethane	0.384	0.406	-5.7	88	0.00
61 T	1,2-Dibromoethane	0.408	0.406	0.5	86	0.00
62 S	4-Bromofluorobenzene	0.452	0.491	-8.6	95	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	90	0.00
64 T	Tetrachloroethene	0.341	0.319	6.5	87	0.00
65 PM	Chlorobenzene	1.079	1.023	5.2	88	0.00
66 T	1,1,1,2-Tetrachloroethane	0.382	0.387	-1.3	88	0.00
67 C	Ethyl Benzene	1.870	1.826	2.4#	87	0.00
68 T	m/p-Xylenes	0.723	0.711	1.7	87	0.00
69 T	o-Xylene	0.725	0.703	3.0	87	0.00
70 T	Styrene	1.163	1.194	-2.7	89	0.00
71 P	Bromoform	0.302	0.320	-6.0	90	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	105	0.00
73 T	Isopropylbenzene	3.668	3.185	13.2	89	0.00
74 T	N-amyl acetate	1.486	1.259	15.3	85	0.00
75 P	1,1,2,2-Tetrachloroethane	1.290	1.120	13.2	92	0.00
76 T	1,2,3-Trichloropropane	1.156	0.986	14.7	88	0.00
77 T	Bromobenzene	0.908	0.800	11.9	92	0.00
78 T	n-propylbenzene	4.143	3.836	7.4	92	0.00
79 T	2-Chlorotoluene	2.554	2.234	12.5	91	0.00
80 T	1,3,5-Trimethylbenzene	3.098	2.782	10.2	90	0.00
81 T	trans-1,4-Dichloro-2-butene	0.361	0.327	9.4	89	0.00
82 T	4-Chlorotoluene	2.876	2.634	8.4	94	0.00
83 T	tert-Butylbenzene	3.131	2.789	10.9	90	0.00
84 T	1,2,4-Trimethylbenzene	3.082	2.792	9.4	92	0.00
85 T	sec-Butylbenzene	3.798	3.575	5.9	94	0.00
86 T	p-Isopropyltoluene	3.165	3.034	4.1	96	0.00
87 T	1,3-Dichlorobenzene	1.692	1.593	5.9	99	0.00
88 T	1,4-Dichlorobenzene	1.732	1.631	5.8	102	0.00
89 T	n-Butylbenzene	2.772	2.887	-4.1	103	0.00
90 T	Hexachloroethane	0.536	0.509	5.0	94	0.00
91 T	1,2-Dichlorobenzene	1.728	1.584	8.3	99	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.257	0.227	11.7	88	0.00
93 T	1,2,4-Trichlorobenzene	1.018	1.084	-6.5	110	0.00
94 T	Hexachlorobutadiene	0.475	0.483	-1.7	105	0.00
95 T	Naphthalene	3.442	3.288	4.5	96	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092623\
Data File : VX037909.D
Acq On : 26 Sep 2023 16:02
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Sep 27 01:16:14 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091523W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 16 04:55:44 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.047	1.034	1.2	103	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6