



BFB

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092623\
Data File : VX037897.D
Acq On : 26 Sep 2023 08:26
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091523W.M
Title : SW846 8260
Last Update : Sat Sep 16 04:55:44 2023



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	17.8	7568	PASS
75	95	30	60	49.3	20931	PASS
95	95	100	100	100.0	42416	PASS
96	95	5	9	6.5	2772	PASS
173	174	0.00	2	0.9	314	PASS
174	95	50	100	79.8	33851	PASS
175	174	5	9	7.0	2386	PASS
176	174	95	101	95.8	32440	PASS
177	176	5	9	6.2	2025	PASS