

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092624\
 Data File : VX043166.D
 Acq On : 26 Sep 2024 10:30
 Operator : JC/MD
 Sample : VX0926WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0926WBL01

Quant Time: Sep 27 01:38:35 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092324W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 24 03:13:52 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	83569	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	165873	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	156853	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	80846	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	74683	53.802	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 107.600%	
35) Dibromofluoromethane	5.385	113	58035	49.597	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 99.200%	
50) Toluene-d8	8.647	98	203930	51.513	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 103.020%	
62) 4-Bromofluorobenzene	11.079	95	83572	52.421	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 104.840%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092624\
Data File : VX043166.D
Acq On : 26 Sep 2024 10:30
Operator : JC/MD
Sample : VX0926WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0926WBL01

Quant Time: Sep 27 01:38:35 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092324W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 24 03:13:52 2024
Response via : Initial Calibration

