

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092624\
 Data File : VX043171.D
 Acq On : 26 Sep 2024 12:29
 Operator : JC/MD
 Sample : P4164-13
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE135D1-20240920

Quant Time: Sep 27 01:40:22 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092324W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 24 03:13:52 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	82959	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	168482	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	152978	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	71334	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	75181	54.559	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 109.120%	
35) Dibromofluoromethane	5.385	113	58184	48.955	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 97.900%	
50) Toluene-d8	8.647	98	205284	51.052	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 102.100%	
62) 4-Bromofluorobenzene	11.079	95	77030	47.569	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 95.140%	
Target Compounds						
					Qvalue	
30) Chloroform	5.098	83	3339	1.713	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092624\
Data File : VX043171.D
Acq On : 26 Sep 2024 12:29
Operator : JC/MD
Sample : P4164-13
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE135D1-20240920

Quant Time: Sep 27 01:40:22 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092324W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 24 03:13:52 2024
Response via : Initial Calibration

