

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092624\
 Data File : VX043172.D
 Acq On : 26 Sep 2024 12:52
 Operator : JC/MD
 Sample : P4164-18
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TT180D-20240923

Quant Time: Sep 27 01:40:38 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092324W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 24 03:13:52 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	89003	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	174450	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	159644	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	70373	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	77305	52.290	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 104.580%	
35) Dibromofluoromethane	5.385	113	59893	48.669	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 97.340%	
50) Toluene-d8	8.647	98	212480	51.034	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 102.060%	
62) 4-Bromofluorobenzene	11.079	95	79325	47.310	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 94.620%	
Target Compounds						
44) Trichloroethene	7.129	130	2712	2.221	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092624\
Data File : VX043172.D
Acq On : 26 Sep 2024 12:52
Operator : JC/MD
Sample : P4164-18
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TT180D-20240923

Quant Time: Sep 27 01:40:38 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092324W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 24 03:13:52 2024
Response via : Initial Calibration

