

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092624\
 Data File : VX043174.D
 Acq On : 26 Sep 2024 13:38
 Operator : JC/MD
 Sample : P4204-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-RW5FRACTANKSED-20240925

Quant Time: Sep 27 01:41:06 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092324W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 24 03:13:52 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	98761	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	194915	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	167295	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	74779	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	89020	54.265	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	108.540%
35) Dibromofluoromethane	5.385	113	67040	48.757	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.520%
50) Toluene-d8	8.647	98	229398	49.313	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.620%
62) 4-Bromofluorobenzene	11.079	95	83163	44.392	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	88.780%
Target Compounds						
16) Acetone	2.386	43	4258	5.550	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092624\
Data File : VX043174.D
Acq On : 26 Sep 2024 13:38
Operator : JC/MD
Sample : P4204-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-RW5FRACTANKSED-20240925

Quant Time: Sep 27 01:41:06 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092324W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 24 03:13:52 2024
Response via : Initial Calibration

