

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092624\
 Data File : VX043164.D
 Acq On : 26 Sep 2024 09:40
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Sep 27 01:37:43 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092324W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 24 03:13:52 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	82	0.00
2 T	Dichlorodifluoromethane	0.533	0.477	10.5	75	0.00
3 P	Chloromethane	0.544	0.502	7.7	81	0.00
4 C	Vinyl Chloride	0.553	0.522	5.6#	81	0.00
5 T	Bromomethane	0.318	0.317	0.3	71	0.00
6 T	Chloroethane	0.354	0.349	1.4	85	0.00
7 T	Trichlorofluoromethane	0.982	0.911	7.2	79	0.00
8 T	Diethyl Ether	0.330	0.366	-10.9	94	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.554	0.549	0.9	85	0.00
10 T	Methyl Iodide	0.658	0.673	-2.3	84	0.00
11 T	Tert butyl alcohol	0.185	0.200	-8.1	92	0.00
12 CM	1,1-Dichloroethene	0.530	0.515	2.8#	83	0.00
13 T	Acrolein	0.024	0.025	-4.2	133	0.00
14 T	Allyl chloride	0.984	0.993	-0.9	89	0.00
15 T	Acrylonitrile	0.322	0.368	-14.3	98	0.00
16 T	Acetone	0.388	0.376	3.1	78	0.00
17 T	Carbon Disulfide	1.317	1.153	12.5	77	0.00
18 T	Methyl Acetate	1.045	1.282	-22.7	98	0.00
19 T	Methyl tert-butyl Ether	2.019	2.234	-10.6	95	0.00
20 T	Methylene Chloride	0.618	0.639	-3.4	92	0.00
21 T	trans-1,2-Dichloroethene	0.548	0.546	0.4	88	0.00
22 T	Diisopropyl ether	1.842	2.034	-10.4	96	0.00
23 T	Vinyl Acetate	2.167	2.393	-10.4	96	0.00
24 P	1,1-Dichloroethane	1.065	1.130	-6.1	91	0.00
25 T	2-Butanone	0.501	0.540	-7.8	91	0.00
26 T	2,2-Dichloropropane	1.078	1.034	4.1	84	0.00
27 T	cis-1,2-Dichloroethene	0.683	0.728	-6.6	92	0.00
28 T	Bromochloromethane	0.469	0.533	-13.6	101	0.00
29 T	Tetrahydrofuran	0.301	0.329	-9.3	95	0.00
30 C	Chloroform	1.175	1.237	-5.3#	91	0.00
31 T	Cyclohexane	0.817	0.763	6.6	80	0.00
32 T	1,1,1-Trichloroethane	1.113	1.072	3.7	83	0.00
33 S	1,2-Dichloroethane-d4	0.831	0.846	-1.8	87	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	86	0.00
35 S	Dibromofluoromethane	0.353	0.345	2.3	86	0.00
36 T	1,1-Dichloropropene	0.433	0.407	6.0	86	0.00
37 T	Ethyl Acetate	0.544	0.557	-2.4	97	0.00
38 T	Carbon Tetrachloride	0.558	0.490	12.2	81	0.00
39 T	Methylcyclohexane	0.544	0.480	11.8	81	0.00
40 TM	Benzene	1.279	1.286	-0.5	90	0.00
41 T	Methacrylonitrile	0.271	0.310	-14.4	97	0.00
42 TM	1,2-Dichloroethane	0.543	0.562	-3.5	93	0.00
43 T	Isopropyl Acetate	0.909	0.941	-3.5	94	0.00
44 TM	Trichloroethene	0.350	0.336	4.0	86	0.00
45 C	1,2-Dichloropropane	0.319	0.332	-4.1#	93	0.00
46 T	Dibromomethane	0.268	0.281	-4.9	96	0.00
47 T	Bromodichloromethane	0.558	0.581	-4.1	92	0.00
48 T	Methyl methacrylate	0.447	0.468	-4.7	97	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092624\
 Data File : VX043164.D
 Acq On : 26 Sep 2024 09:40
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleID :
 VSTDCCC050

Quant Time: Sep 27 01:37:43 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092324W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 24 03:13:52 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.008	0.009	-12.5	95	0.00
50 S	Toluene-d8	1.193	1.136	4.8	84	0.00
51 T	4-Methyl-2-Pentanone	0.559	0.596	-6.6	96	0.00
52 CM	Toluene	0.824	0.828	-0.5#	89	0.00
53 T	t-1,3-Dichloropropene	0.573	0.593	-3.5	92	0.00
54 T	cis-1,3-Dichloropropene	0.572	0.599	-4.7	93	0.00
55 T	1,1,2-Trichloroethane	0.339	0.360	-6.2	95	0.00
56 T	Ethyl methacrylate	0.590	0.628	-6.4	95	0.00
57 T	1,3-Dichloropropane	0.577	0.606	-5.0	96	0.00
58 T	2-Chloroethyl Vinyl ether	0.239	0.249	-4.2	89	0.00
59 T	2-Hexanone	0.437	0.461	-5.5	92	0.00
60 T	Dibromochloromethane	0.443	0.462	-4.3	94	0.00
61 T	1,2-Dibromoethane	0.371	0.393	-5.9	96	0.00
62 S	4-Bromofluorobenzene	0.481	0.465	3.3	85	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	87	0.00
64 T	Tetrachloroethene	0.362	0.326	9.9	82	0.00
65 PM	Chlorobenzene	1.079	1.075	0.4	91	0.00
66 T	1,1,1,2-Tetrachloroethane	0.398	0.398	0.0	91	0.00
67 C	Ethyl Benzene	1.866	1.821	2.4#	88	0.00
68 T	m/p-Xylenes	0.694	0.688	0.9	89	0.00
69 T	o-Xylene	0.696	0.684	1.7	90	0.00
70 T	Styrene	1.162	1.216	-4.6	92	0.00
71 P	Bromoform	0.350	0.358	-2.3	90	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	86	0.00
73 T	Isopropylbenzene	3.750	3.558	5.1	87	0.00
74 T	N-amyl acetate	1.762	1.827	-3.7	93	0.00
75 P	1,1,2,2-Tetrachloroethane	1.207	1.251	-3.6	94	0.00
76 T	1,2,3-Trichloropropane	1.063	1.113	-4.7	94	0.00
77 T	Bromobenzene	0.900	0.917	-1.9	92	0.00
78 T	n-propylbenzene	4.254	4.166	2.1	88	0.00
79 T	2-Chlorotoluene	2.668	2.555	4.2	88	0.00
80 T	1,3,5-Trimethylbenzene	3.102	2.998	3.4	87	0.00
81 T	trans-1,4-Dichloro-2-butene	0.446	0.448	-0.4	91	0.00
82 T	4-Chlorotoluene	3.050	3.020	1.0	89	0.00
83 T	tert-Butylbenzene	3.197	3.044	4.8	88	0.00
84 T	1,2,4-Trimethylbenzene	3.087	3.044	1.4	88	0.00
85 T	sec-Butylbenzene	3.929	3.691	6.1	85	0.00
86 T	p-Isopropyltoluene	3.262	3.132	4.0	86	0.00
87 T	1,3-Dichlorobenzene	1.641	1.658	-1.0	91	0.00
88 T	1,4-Dichlorobenzene	1.646	1.658	-0.7	92	0.00
89 T	n-Butylbenzene	2.843	2.831	0.4	87	0.00
90 T	Hexachloroethane	0.650	0.621	4.5	84	0.00
91 T	1,2-Dichlorobenzene	1.616	1.637	-1.3	93	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.326	0.319	2.1	89	0.00
93 T	1,2,4-Trichlorobenzene	1.002	1.041	-3.9	93	0.00
94 T	Hexachlorobutadiene	0.421	0.376	10.7	81	0.00
95 T	Naphthalene	3.483	3.671	-5.4	95	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092624\
Data File : VX043164.D
Acq On : 26 Sep 2024 09:40
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Sep 27 01:37:43 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092324W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 24 03:13:52 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.996	1.036	-4.0	94	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6