

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092624\
 Data File : VX043166.D
 Acq On : 26 Sep 2024 10:30
 Operator : JC/MD
 Sample : VX0926WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0926WBL01

Quant Time: Sep 27 01:38:35 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092324W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 24 03:13:52 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	83569	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	165873	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	156853	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	80846	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	74683	53.802	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	107.600%
35) Dibromofluoromethane	5.385	113	58035	49.597	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.200%
50) Toluene-d8	8.647	98	203930	51.513	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	103.020%
62) 4-Bromofluorobenzene	11.079	95	83572	52.421	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	104.840%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092624\
Data File : VX043166.D
Acq On : 26 Sep 2024 10:30
Operator : JC/MD
Sample : VX0926WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0926WBL01


Quant Time: Sep 27 01:38:35 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092324W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 24 03:13:52 2024
Response via : Initial Calibration








#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.550 min Scan# 719
Delta R.T. -0.000 min
Lab File: VX043166.D
Acq: 26 Sep 2024 10:30

Instrument :
MSVOA_X
ClientSampleId :
VX0926WBL01







Tgt Ion:168 Resp: 83569
Ion Ratio Lower Upper
168 100
99 66.2 50.8 76.2



#33
1,2-Dichloroethane-d4
Concen: 53.802 ug/l
RT: 5.958 min Scan# 799
Delta R.T. 0.006 min
Lab File: VX043166.D
Acq: 26 Sep 2024 10:30

Tgt Ion: 65 Resp: 74683
Ion Ratio Lower Upper
65 100
67 49.8 0.0 99.0









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.763 min Scan# 91
Delta R.T. 0.006 min
Lab File: VX043166.D
Acq: 26 Sep 2024 10:30




Instrument :
MSVOA_X
ClientSampleId :
VX0926WBL01


Tgt Ion:114 Resp: 165873
Ion Ratio Lower Upper
114 100
63 21.2 0.0 43.2
88 17.2 0.0 33.4



#35
Dibromofluoromethane
Concen: 49.597 ug/l
RT: 5.385 min Scan# 705
Delta R.T. 0.006 min
Lab File: VX043166.D
Acq: 26 Sep 2024 10:30

Tgt Ion:113 Resp: 58035
Ion Ratio Lower Upper
113 100
111 101.2 82.6 124.0
192 17.6 14.6 22.0









#50
Toluene-d8
Concen: 51.513 ug/l
RT: 8.647 min Scan# 1133
Delta R.T. -0.000 min
Lab File: VX043166.D
Acq: 26 Sep 2024 10:30




Instrument :
MSVOA_X
ClientSampleId :
VX0926WBL01


Tgt Ion: 98 Resp: 203930
Ion Ratio Lower Upper
98 100
100 65.2 52.2 78.4



#62
4-Bromofluorobenzene
Concen: 52.421 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. -0.000 min
Lab File: VX043166.D
Acq: 26 Sep 2024 10:30

Tgt Ion: 95 Resp: 83572
Ion Ratio Lower Upper
95 100
174 72.2 0.0 146.0
176 70.5 0.0 142.6









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. -0.000 min
Lab File: VX043166.D
Acq: 26 Sep 2024 10:30

Instrument :
MSVOA_X
ClientSampleId :
VX0926WBL01

Tgt Ion:117 Resp: 156853
Ion Ratio Lower Upper
117 100
82 57.4 44.7 67.1
119 32.5 26.3 39.5



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.024 min Scan# 1794
Delta R.T. 0.006 min
Lab File: VX043166.D
Acq: 26 Sep 2024 10:30

Tgt Ion:152 Resp: 80846
Ion Ratio Lower Upper
152 100
115 63.7 43.7 131.1
150 158.0 0.0 344.6

