

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092625\
 Data File : VX047850.D
 Acq On : 26 Sep 2025 11:43
 Operator : JC/MD
 Sample : VX0926WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0926WBL01

Quant Time: Sep 26 23:11:24 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	182475	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.745	114	379369	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	380700	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	183902	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	146381	50.397	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	100.800%
35) Dibromofluoromethane	5.373	113	123091	48.380	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.760%
50) Toluene-d8	8.635	98	433662	49.259	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.520%
62) 4-Bromofluorobenzene	11.061	95	177814	53.220	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	106.440%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092625\
Data File : VX047850.D
Acq On : 26 Sep 2025 11:43
Operator : JC/MD
Sample : VX0926WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0926WBL01

Quant Time: Sep 26 23:11:24 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

